



6375 Dixie Rd Unit# 7,
Mississauga, ON L5T 2E7
Tel: (905)565-1583
Fax: (905)565-8325

Form:R-CalKit	Approved By:JR Jun/17	Ver1.0
---------------	-----------------------	--------

150038

Make	Model	Serial	Asset	By
Siborg Systems	R-Cal Kit	1	1	JR

Resistance accuracy at DCV

Position	Nom Value		Measured	Error %
1	499	kΩ	499.422	0.085
2	75	kΩ	74.995	-0.007
3	30	kΩ	30.001	0.003
4	18	kΩ	17.997	-0.016
5	7.5	kΩ	7.4994	-0.008
6	3.6	kΩ	3.5985	-0.042
7	1.24	kΩ	1.2396	-0.031
8	300	Ω	300.082	0.027
9	150	Ω	149.99	-0.007
10	51	Ω	50.972	-0.054
11	30	Ω	29.957	-0.142
12	15	Ω	14.957	-0.287
13	4.7	Ω	4.66	-0.921
14	0	Ω	Reference	